

(A) **TECHNICAL SPECIFICATION FOR BED**

(1) **FOWLER DELUXE BED**

- (a) Frame work of strong and precise steel rectangular tubes of **50x30 mm (16 gauge) mounted on 10 CMS Dia swiveling castors wheels** (two with brakes).
- (b) Having four section top **perforated 16 gauge CRCA sheet** duly machine press bent with mattress guards.
- (c) Backrest and kneerest sections maneuvered by two separate screws from foot end.
- (d) SBR moulded head and foot bows.
- (e) Driving **handle should be made of ABS Injection moulded**, with position indication, foldable and slideble "In" & "Out".
- (f) Size : **203L x 90W x 60 H Cms.**
- (g) Provision for Four I.V. Rod Locations.
- (h) Provision of installing a mosquito net rods.
- (i) Provision of attaching paper roll.
- (j) Provision of clip for Mattress and Bed Sheet so that mattress does not slip.
- (k) Provision of side railing which should be foldable and slidable under the bed.
- (l) FINISH : Antirust treatment and Epoxy powder coated.

(2) **FOWLER BED**

- (a) Frame work of strong and precise steel rectangular tubes of **50x30 mm (16 gauge) mounted on 10 CMS Dia swiveling castors wheels** (two with brakes).
- (b) Having four section top **perforated 16 gauge CRCA sheet** duly machine press bent with mattress guards.
- (c) Backrest and kneerest sections maneuvered by **two separate screws** from foot end.
- (d) Tabular head and foot bows of unequal **height of 16 gauge** with vertical tube supports.
- (e) Handle ABS moulded. Foldable and slideble in and out.
- (f) Size : **203L x 90W x 60 H Cms.**
- (g) Provision for Four I.V. Rod Locations.
- (h) Provision of installing a mosquito net rods.
- (i) Provision of attaching paper roll.
- (j) Provision of clip for Mattress and Bed Sheet so that Mattress does not slip.
- (k) Provision of stainless steel side railing which should be foldable and slidable under the bed.
- (l) FINISH : Antirust treatment and Epoxy powder coated.

(3) SEMI FOWLER BED

- (a) Frame work of strong and precise steel rectangular tubes of **50x30 mm (16 gauge) mounted on 10 CMS Dia swiveling castors** wheels (two with brakes).
- (b) Having two section top **perforated 16 gauge CRCA sheet** duly machine press bent with mattress guards.
- (c) Backrest and kneerest sections maneuvered by two separate screws from foot end.
- (d) Tabular head and foot bows of unequal **height of 16 gauge** with vertical tube supports.
- (e) **Handle ABS moulded. Foldable and slideble** in and out.
- (f) **Size : 203L x 90W x 60 H Cms.**
- (g) Provision for Four I.V. Rod Locations.
- (h) Provision of installing a mosquito net rods.
- (i) Provision of attaching paper roll.
- (j) Provision of clip for Mattress and Bed Sheet so that Mattress does not slip.
- (k) Provision of stainless steel side railing which should be foldable and slidable under the bed.
- (l) **FINISH : Antirust treatment and Epoxy powder coated.**