

State Health Society, Bihar
Pariwar Kalyan Bhawan, Sheikhpura, Patna-14

EQUIPMENT-WISE EVALUATION OF BIDS RECEIVED AGAINST TENDER
REF NO. SHSB/2012-13/RAD-PATH- 001

Equipment Name	Remarks
Ultrasound machine	The equipment quoted by South India Surgical Co., and by Kopran Laboratories, does not have FDA certificate, hence their bid is rejected, Hospital Supply Co. has provided EMD in the form of Fixed Deposit Certificate with validity till 26 th Feb 2013, which does not satisfy 150 days validity as required and S S Medical Systems provided EMD in the form of cheque, hence its bid is rejected. The equipment will be retendered.
Color Doppler	The equipment quoted by South India Surgical Co. does not have FDA certificate, hence its bid is rejected, Hospital Supply Co. has provided EMD in the form of Fixed Deposit Certificate with validity till 26 th Feb 2013, which does not satisfy 150 days validity as required and S S Medical Systems provided EMD in the form of cheque, hence its bid is rejected. Since Philips is the only bidder left, the equipment will be retendered.
X ray 300mA	General Medical Equipment has quoted LCD display in place of LED display, hence its bid is technically rejected. Capability of ME X ray has anode speed of 2700 rpm in place of 2800 rpm hence its bid is rejected. Since Allengers is the only bidder left, the equipment will be retendered.
X ray 500 mA	LED readout parameter is not provided by GME, Allengers and ME X ray system their bids are rejected. Since Nucleotech is the only bidder left, the equipment will be retendered.
Semi Auto Analyser	Robonik does not match technical specifications, Model quoted by Priman Instruments is filter based and that required in Technical specifications is grating based and S. S. Medical has submitted EMD in the form of cheque, hence their bids are rejected. Since Kopran is the only bidder left, the equipment will be retendered.
Photoelectric Colorimeter	Both the bidders Yorco Slaes and Priman does not satisfy 80% supply criteria, hence their bids are rejected. The equipment will be retendered.

It is also decided that since it is not justified to procure CR System without procuring X Ray machines, CR System will also be retendered.

[Handwritten Signature]
14/03/13

Jai Prakash Singh (BAS)
Senior Deputy Collector

[Handwritten Signature]
Executive Director
SHSB, Patna