



राज्य स्वास्थ्य सभिति, बिहार



GOVERNMENT OF BIHAR.
DEPARTMENT OF HEALTH

e-tender (NIT) Reference No.: - 06/SHSB/PPP(DIALYSIS)/2019-20

CORRIGENDUM

1. With reference to the **Notice Inviting Tender (NIT) Reference No.: - 06/SHSB/PPP(DIALYSIS)/2019-20** published in different newspapers by PR No.- 005635 (Ni.Ni) 2019-20 and uploaded on the website " <http://www.eproc.bihar.gov.in/BELTRON> ", for setting up and operationalizing 24X7 dialysis units through Public Private Partnership (PPP) mode in 38 government hospitals in the state of Bihar, the Pre-Bid meeting was held at **11:00 AM on 08.08.2019** in the Conference Hall of the State Health Society, Bihar (SHSB), Pariwar Kalyan Bhawan, Sheikhpura, Patna.

2. In response to the queries, raised during the Pre-bid meeting and submitted via email, the Technical Committee after due consideration, decides to the following Corrigendum/Amendments in the RFP:

S. No.	Request by the bidder	Reference in the Tender Document	Relevant clause as published in the Tender	Corrigendum/Amendments
1	Amendments to the technical specifications of Haemodialysis machine	Appendix L: Haemodialysis machine & Associated systems, Clause H, Point 1.4.1, under Technical Specification of Haemodialysis machine Page 56 of 68. And Appendix L: Haemodialysis machine & Associated systems, Clause H, Point 1.4.3, under Technical Specification of Haemodialysis machine Page 56 of 68.	1.4.1 Venous Pressure Monitoring Range should be approx <u>700 to 750 mm Hg</u> accuracy +- 5 mm Hg or +- 10% whichever is greater. And 1.4.3 Arterial Pressure Monitoring – It should be approx <u>700 to 750 mm Hg</u> .	<u>Amendments:-</u> 1.4.1. Venous Pressure Monitoring Range should be approx (-) <u>700 to 750 mm Hg</u> , accuracy +- 5 mm Hg or +- 10% whichever is greater. And 1.4.3 Arterial Pressure Monitoring – It should be approx (-) <u>700 to 750 mm Hg</u> .

S. No.	Request by the bidder	Reference in the Tender Document	Relevant clause as published in the Tender	Corrigendum/Amendments
2	Amendments to the technical specifications of the Water Treatment Plant	Appendix L: Haemodialysis machine & Associated systems, Clause H, sub-clause 3, Point 3.1, Page 59 of 68.	3.11 The system should comprise of pre-treatment modules such as sand filter, activated carbon filter, water softener, 5-micron particulate filter and <u>deionizer</u> before the reverse osmosis unit and post R.O. Bacterial Filters (1 micron) and UV light Disinfection for yielding high purity water.	<u>Amendments:-</u> 3.1 The system should comprise of pre-treatment modules such as sand filter, activated carbon filter, water softener and 5-micron particulate filter before the reverse osmosis unit and post R.O. Bacterial Filters (1 micron) and UV light Disinfection for yielding high purity water.
3	Amendments to the technical specifications of the Water Treatment Plant	Appendix L: Haemodialysis machine & Associated systems, Clause H, sub-clause 3, Point 3.11, Page 60 of 68.	3.11 The water distribution loop, booster pump and storage water tank should be made up of <u>PVC</u> (High quality), Storage water tank should have capacity of at least 2,000 litres with water level controller, outlet valves and easy cleaning provisions.	<u>Amendments:-</u> 3.11 The water distribution loop, booster pump and storage water tank should be made up of <u>PEX</u> (High quality), Storage water tank should have capacity of at least 2,000 litres with water level controller, outlet valves and easy cleaning provisions.

3. In view of the changes incorporated in the tender documents as per the queries received in the Pre-Bid Meeting, it has been decided to amend the dates, and therefore the revised Schedule of Events is mentioned below:

S. No	Event Description	Timeline
1	Date & time of downloading the RFP	Till 05/09/2019 (Thursday) up to 05:00 PM, on the e-Procurement Portal (https://www.eproc.bihar.gov.in/BELTRON).
2	Last date & time for submission (upload) of online bidding document	06/09/2019 (Friday) till 05:00 PM, on the e-Procurement Portal (https://www.eproc.bihar.gov.in/BELTRON)



S. No	Event Description	Timeline
3	Last date & time for submission of EMD in Hard (Physical) Copies (Offline Mode)	12/09/2019 (Thursday) by 05:00 PM, To, "The Executive Director, State Health Society, Bihar, Pariwar Kalyan Bhawan, Sheikhpura, Patna-800014"
4	Time, Date of opening of Technical Bid	13/09/2019 (Friday) at 11:00 AM on the e-Procurement Portal (https:// www.eproc.bihar.gov.in / BELTRON
5	Time, Date of opening of Financial Bid	To be announced later on the e-Procurement Portal (https://www.eproc.bihar.gov.in/BELTRON)

4. All amendments in the RFP, as mentioned above are applicable to the bidders.

5. Other terms and conditions of the RFP shall remain the same.


 Executive Director
 State Health society Bihar